



EcoChyll X1 Rotary Evaporator



OPERATING MANUAL

ECOCHYLL S SUMMARY DESCRIPTION

Rotary evaporators (rotovaps) are ubiquitous devices in chemistry labs and industries performing chemistry, including labs in the chemical, pharmaceutical, environmental, materials, life science, and cannabis industries. Rotovaps consist of a heating fluid bath, rotating motor, evaporating flask, condenser, collection flask, and vacuum source. Traditional rotovap condensers require external sources of cooling materials such as dry ice, liquid nitrogen, water or glycol. Glycol requires additional recirculating chiller equipment, which are often bulky, heavy, and inefficient. The traditional rotovap for decades have also been characterized with inefficient vapor condensing, operational cost, unreliability, high energy consumption and material waste.

Using a proprietary and innovative self-cooling technology, Ecodyst has revolutionized the rotovap to be more efficient, to have a smaller footprint, to have greater output, and to be less costly to operate as compared with traditional rotovaps. This disruptive technology has set new standards worldwide for rotary evaporators.

Ecodyst's green condenser technology does not require an external source of cooling materials, eliminating the major sources of material waste associated with rotovaps. Ecodyst's devices are equipped with intelligent self-cooling condenser technology that is extremely efficient at condensing vapors. This technology is environmentally friendly, energy efficient, reliable and sustainable. Additionally, the condenser achieves rapid cool down (reaches -40oC in 60 seconds) and is always-available, which eliminates downtime, increases productivity, and saves time and money.

EcoChyll X1 is a powerful, small footprint and intelligent self-cooling condenser invented for scientists by scientists. Additionally, this condenser has superior advantages, has a large cooling survey area condenser and extremely fast rates of evaporation. The EcoChyll S is a plug-and-play condenser that upgrades any bench top rotovap up to 5-Liter. EcoChyll eliminates the hassle associated with rotovap condensers and free up time for scientists to focus on more complex tasks.

Superior advantages of the EcoChyll X1 Rotovap

1. Intelligent self-cooling condenser
2. Fast rates of evaporation
3. Small footprint
4. Always-available condenser
5. No dry ice or external chiller
6. No associated operational cost
7. High levels of productivity
8. Preserves water, no water bills
9. Environmentally friendly
10. Pays for itself within 3 to 5 years

ABOUT THIS MANUAL

This operating manual provides the technical and operational details for the EcoChyll X1.

- Please read this manual carefully and obey all safety and warning notices.
- Ensure that every operator reads this manual.
- Ensure that this manual is accessible for every operator.
- Pass on the operating manual to the subsequent owner.
- In addition, please observe the regional regulations

TABLE OF CONTENTS

***SAFETY INSTRUCTIONS*..... 5**

 General Safety Instructions5

 Intended use5

 Improper use5

 Installation / Electrical Safety.....5

 Personnel Qualification6

 Operating Company’s Obligations6

 Safety During Use7

 Disposal.....7

***ACCESSORIES AND SPARE PARTS* 8**

 Scope of Delivery8

***DEVICE DESCRIPTION* 9**

 Device Overview: ECOCHYLL X1 CONNECTS TO ANY BENCHTOP ROTOVAP9

 EcoChyll X1 Controller9

 Technical Data10

***ASSEMBLY*.....10**

 Installation.....11

***POST OPERATIONS*.....12**

 Cleaning.....12

 Maintenance12

 Troubleshooting.....12

 Transportation and Storage.....12

***SUPPORT*.....12**

 Warranty12

***SERVICE*13**

 Confirmation of Condition of Unit13

***CONTACT DETAILS*14**

SAFETY INSTRUCTIONS

General Safety Instructions

The device has been constructed according to state-of-the-art technology and recognized safety regulations. However, risks may still arise during installation, operation and maintenance.

Please ensure the operating manual is available at all times.

The device may only be used under the following circumstances

- Only operate the device, if it is in full working order.
- Ensure all operators of the device possess the necessary safety and risk awareness.
- Operate the device according to the instructions stipulated in this manual only.
- If there is something you do not understand, or certain information is missing, ask your manager or contact the manufacturer.
- Do not do anything on the device without authorization.
- Only use the device according to its intended use.

Intended use

The device is intended for use by trained and authorized personnel only. The device is suitable for the following

- Economical cooling as alternative to tap water systems use
- Research application

Improper use

Any use which deviates from the device's intended use is considered to be improper. The manufacturer does not accept liability for any damages resulting from non-permitted uses. The risk is carried by the operator alone.

Installation / Electrical Safety

- The device may only be connected when the mains voltage corresponds to the information on the type plate of the unit.
- The mains connection must be accessible at all times.
- Repairs may only be performed by a qualified electrician.
- Never operate the unit with a damaged power cord.
- Always turn the unit OFF and disconnect mains power before performing any maintenance or service.

Personnel Qualification

- The device may only be operated by qualified persons.
- The device may only be operated by individuals who have been instructed in its proper use by qualified persons.
- The device may only be operated and maintained by persons who are of legal age.
- Other personnel may only work with the unit under continuous supervision of an experienced qualified operator.
- This manual must be read and understood by all persons working with the device.
- The personnel must have received special safety instructions in order to guarantee responsible and safe work procedures.

Operating Company's Obligations

A. Installation Site

- The device must be positioned in a suitable location.
- The device must be installed sufficiently stable on a strong and level surface.
- All screw connections must be securely tightened.
- The device should be located as close as possible to the process requiring cooling.
- The device should not be installed closer than 1.4 meters (4 feet) to a heat-generating source, such as heating pipes, boilers, etc.
- If possible, the device should be located near a suitable drain to prevent flooding in the event of leaks.
- Do not place the device where corrosive fumes, excessive moisture, excessive dust, or high room temperatures are present.
- The site must have an 80% relative humidity and temperature between 5 - 35 °C
- Adequate clearance should be allowed on the front, sides, top, and rear of the device for access to connections and components.
- The front and side vents of the device must be a minimum of 21 cm (8 inches) away from walls or vertical surfaces, so air flow is not restricted.

B. Changes to the Unit

- No unauthorized changes may be made to the unit.
- No parts may be used which have not been approved by the manufacturer.
- Unauthorized changes result in the EC Declaration of Conformity losing its validity, and the appliance may no longer be operated.
- The manufacturer is not liable for any damage, danger or injuries that result from unauthorized changes or from operating the unit other than described in this manual.

C. Safety for the Personnel

Ensure that only qualified personnel operate the device. Observe the following regulations:

- Laboratory guidelines
- Accident prevention regulation
- Ordinance on Hazardous Substances
- Other generally accepted rules of safety engineering and occupational health
- Local regulations

Safety During Use








- Wear the appropriate protective clothing when working on the device (clothing, protective glasses and if necessary, safety gloves).
- Do not use the device in potentially explosive areas. The device is not protected against explosion. There is no explosion or ATEX protection available.
- Do not operate or assemble devices in the vicinity which are emission or radiation sources (electromagnetic waves) for the frequency range (3×10^{11} Hz to 3×10^{15} Hz).
- Avoid putting pressure on the display when you do not operate the device. • Eliminate errors immediately.
- Do not use abrasive material to clean the glass surfaces. Only wipe with damp cloths.
- Always switch the device off after use.

Disposal

- Check the device components for hazardous substances and solvents.
- Clean all components before disposal.
- Dispose of the device according to the valid national legal regulations.
- Dispose of the packaging material in accordance with the valid national legal regulations. • Have refrigerant emptied before disposal

ACCESSORIES AND SPARE PARTS

Scope of Delivery

Qty	Part No		Description
1	ECO-007-X1		EcoChyll X1 115V Complete with Condenser Body, To Include:
1	ECO-777-091		Condenser Glass Body, 100mm O-Ring Groove Flange
1	ECO-777-155		Adapter for connecting Rotovap to EcoChyll
1	ECO-777-166		100mm Quick Release Clamp
1	ECO-777-164		100mm Viton O-Ring
1	ECO-777-007		Vapor and Vacuum Tubing 5 ft
1	ECO-777-001		1000ml Round Bottom Flask, 24/40 Outer Joint
1	ECO-777-002		1000ml Receiving Flask, 35/25 Ball Joint



DEVICE DESCRIPTION

Device Overview: ECOCHYLL X1 CONNECTS TO ANY BENCHTOP ROTOVAP



EcoChyll X1 Controller



Press  or  to set the value as required.

Technical Data

Standard Supply Voltage	115V, 15-amp, 60 Hz
Power Supply (W)	656
Default Set Temperature	-40
Temp stability of cold finger	+/- 0.1
Operating Temperature Range (C)	-5....-40
Weight (lbs)	85
Material	Glass condenser, chemically resistant metallic alloy condenser, powder coated sheet metal body
Side Vents	Vents must be unblocked for proper cooling transfer
Dimensions (in) (W x D x H)	12 x 17.5 x 27.25
Shipping weight (lbs)	175
Shipping Dimensions (in)	26.5 x 22.5 x 34.5

EcoChyll X1 Cooling Capacity Specs

Evap. temp		Capacity	Power cons.	Current cons.	EER
°F	o C	BTU/h +/-5%	W +/-5%	A +/-5%	BTU/Wh
-40	-40	122	145	4.21	0.84
-30	-34	482	248	4.64	1.95
-20	-29	909	339	5.10	2.68
-13	-25	1257	398	5.44	3.16
-10	-23	1420	422	5.58	3.36
0	-18	2034	500	6.09	4.07
10	-12	2769	573	6.63	4.83
14	-10	3101	602	6.85	5.15
20	-7	3643	645	7.19	5.64
30	-1	4677	719	7.78	6.51

ASSEMBLY

Installation

1. Remove EcoChyll from packaging. Save or dispose of packaging material.
2. Place unit in intended workspace.
3. Place O-ring in the condenser groove.
4. Open clamp and put over condenser flange.
5. Slide condenser over Ecochyll condenser.
6. Secure clamp and tighten with a screwdriver.
7. If you have existing rotary evaporator, remove the condenser and replace with the rotovap adapter provided.
8. Put coupling and coupling spring on rotovap adapter and secure to the rotary drive head.
9. Attach one end of supplied tubing to rotovap adapter and the other end to the condenser glass body hose connector. Hose connector is found on bottom side of glassware.
10. Attach receiving flask to condenser with pinch clamp.
11. Plug in Ecochyll and turn the switch to one located on the right-hand side of unit.
12. Ecochyll display will light with 8.8.8.8 for startup period
13. Ecochyll is preset to -37.5C
14. Bottom display bar will show the set temperature and the top display will show the actual temperature of the cold finger. You will begin to see the temperature dropping.
15. Within 1-3 minutes you will see the display reaching lower than -40C
16. Check all ports for vacuum leaks, tighten clamps and adapters as needed. If further leaks occur see recommendations in troubleshooting section.

POST OPERATIONS

Cleaning

Recommended to clean condenser glass after use and replace tubing as needed.

Maintenance

The unit is maintenance-free. Any necessary repair must be performed by an authorized Ecodyst distributor. Please contact Ecodyst or your local Ecodyst distributor

Troubleshooting

Loss of Vacuum	Assess and tighten clamps on various components Check vacuum seal on rotary evaporator
Ice Build Up on condenser	Ensure EcoChyll is not left on for long periods without Condensing

Transportation and Storage

- Store and transport the unit and its components only if they were emptied and cleaned in the original packing material.
- Alternatively store and transport the unit in a container which prevents damages during transportation.
- Seal the packing carefully against unauthorized or mistaken opening.
- Store the unit in a dry place.

SUPPORT

Warranty

Ecodyst provides a 1-year warranty for the products described here (excluding glass and wearing parts) if you register using the warranty card enclosed. The warranty is valid from the point of registration. The serial number is also valid without registering. The warranty covers part and manufacturing defects.

(Serial number on back top corner of EcoChyll S)

Questions / Repair work

If any aspect of installation, operation or maintenance remains unanswered in the present manual, please contact us. For repairs, please call Ecodyst or your local authorized Ecodyst distributor.

SERVICE

Confirmation of Condition of Unit

In the case of repair, copy and complete the Confirmation of condition of unit and send it to Ecodyst.

Product number

Serial number

Reason for repair

Has the device been cleaned, decontaminated/sterilized?

___ Yes ___ No

Is the unit in a condition which does not represent any health threats for the staff of our service department?

___ Yes ___ No

If not, which substances has the unit come into contact with?

Legally binding declaration

The customer is aware of being legally liable to Ecodyst for any damages arising from incomplete and incorrect information.

Signature:

Date:

Please note: The shipper is responsible for the return of the goods in well packed condition, suitable for the mode of transport.

Last name, first name

Company

Street

Zip code, city, country

E-mail

CONTACT DETAILS

Ecodyst, Inc.

1010 Goodworth Dr. Apex, NC 27539

919-717-4061

info@ecodyst.com

www.ecodyst.com



Invented for scientists by scientists