



PERSONAL PRECISION

USER GUIDE




WELCOME

Thank you for choosing the MyGo Mini.

The MyGo Mini is the world's most compact real-time PCR instrument, with no moving parts providing silent operation and maximum reliability. With outstanding precision, it is capable of providing great results over a broad range of applications.

Personal precision...



CONTENT

This quick start guide will teach you what you need to know to start running your MyGo Mini. It describes everything from opening and closing the lid, to data analysis for your qPCR experiments.

Contents include...

SETUP

Learn how to install your MyGo Mini - from setting it up on your workbench, to configuring the instrument and installing the software.

TIPS

Learn some useful tips for using your instrument.

EXPERIMENT

Learn the basics of setting up an experiment on your MyGo Mini.

ANALYSIS

Learn how to easily analyse your results.

EXPORT

Learn how to export your data for use with other software.

MAINTENANCE

Learn how to look after your MyGo Mini.

TROUBLESHOOTING

If you ever have a problem with your instrument, learn how to troubleshoot it here.


SETUP

INSTALLING YOUR MYGO MINI

You should have the following items in the MyGo Mini box:

1. MYGO MINI QPCR INSTRUMENT
2. MYGO MINI USER GUIDE
3. POWER SUPPLY UNIT AND MAINS CABLE
4. ETHERNET CABLE
5. USB DRIVE
6. USB DRIVE EXTENSION CABLE

Please keep the box the MyGo Mini came in, in case you need to protect your MyGo Mini during storage or shipment in future.




LID

The lid of the MyGo Mini can be removed from the instrument and must be attached at all times when not in use. To remove the lid push down and turn anticlockwise. To attach the lid, first line up the lugs on the lid with the grooves on the mount plate as shown below. Once aligned press down, and then turn clockwise to secure until you feel a hard stop.

If the instrument is connected to power the instrument will say "Lid on" and "Lid off" when you attach and remove the heated lid. The instrument will display green when the lid is fitted, and display pulsating yellow when the lid is removed. The lid may get warm during operation; this is normal.

Lugs

Grooves



HEATSINK

This is the outer body of the instrument and enables efficient cooling of the system during thermal cycling. Ensure that there is free airflow around the instrument - do not cover or wrap the instrument, and keep it at least 5 cm away from objects or walls.

FOOT


The foot should always be placed facing down on your worktop during operation.

MYGO MINI CONNECTIONS

If you turn your MyGo Mini upside down you will see three connections as shown below:

POWER

This is to connect your MyGo Mini to the provided power supply unit.



USB PORT

Insert the USB drive here to run an experiment from the USB drive, or to update firmware using a USB drive. You can use the supplied USB drive extension cable.

ETHERNET PORT

This is for connecting your MyGo Mini to your LAN or computer.

POWERING ON YOUR MYGO MINI

Connect the power supply to the MyGo Mini, place the MyGo Mini on your lab bench, and then connect AC power to the power supply unit. Your instrument will now turn on after a few seconds. The display LEDs will light up multi-coloured, and then turn green if a lid is attached, or yellow if not. Your MyGo Mini is ready to run.

CONNECTING YOUR MYGO MINI

TO A NETWORK, PC OR LAPTOP

Connect one end of the ethernet cable to the instrument and the other end to a LAN port, or directly to your computer. Once your MyGo Mini is connected to a LAN, you can connect to your instrument via Wi-Fi if your network supports it.


[A summary of these connections is shown on the following page...](#)




MyGo Mini Connection Modes



1. Direct network connection



2. Local area network connection



3. USB drive connection

MYGO MINI LIGHTS AND SOUNDS

The MyGo Mini uses display LEDs and sounds to help you understand what the instrument is doing. Here is a summary of these display LEDs and sounds:

DISPLAY LEDs

Green	Idle / Ready to run
Yellow	Lid off
Blue	Experiment is starting/ cooling down
Blue	Running, low temperature
Pink	Running, medium temperature
Red	Running, high temperature
Multi-Coloured	Experiment complete
Red-Flashing	Fault detected

SOUNDS

“Lid On”	The lid has been attached
“Lid Off”	The lid has been removed
“Analysis Starting”	An experiment has begun
“Analysis Complete”	An experiment has been completed

INSTALLING YOUR MYGO MINI SOFTWARE


Your MyGo USB drive contains software for Windows, Mac OS X and Linux operating systems. Please open the software file matching your chosen operating system. The latest version of the software can also be downloaded from our website mygopcr.com.

WINDOWS

Double-click on the windows installer and follow the on-screen instructions to install your MyGo Mini software on Windows.


MAC OS X

Double-click on the disk image and follow the on-screen instructions to install your MyGo Mini software on your Mac PC.




CONFIGURING YOUR MYGO MINI


With your MyGo Mini software open and your MyGo Mini connected to the network please connect to your MyGo Mini Instrument.




Please select MyGo Mini



To add your new MyGo Mini select the **Add** button. You will now be presented with a list of available instruments.





Double-click on the instrument you wish to connect to, or press **Select** with the instrument selected. The instrument will now be added to the list of **Registered Instruments** for use in the software.

TIPS

Here are some great tips to keep in mind whilst running your MyGo instrument.

1. LID GETS WARM
2. DO NOT LEAVE THE LID OFF
3. SPIN YOUR TUBES
4. REMOVE ALL BUBBLES
5. KEEP YOUR LAB CLEAN
6. KEEP YOUR INSTRUMENT CLEAN
7. PC SETTINGS

LID GETS WARM

The MyGo Mini lid can get warm during operation, this is completely normal.

DO NOT LEAVE THE LID OFF


If the lid is off dust may fall into the wells and affect the performance of your instrument.

SPIN YOUR TUBES

This should ensure that all contents are at the bottom of the tubes, and there are no bubbles present. Reaction mixtures which are viscous, or contain high levels of detergents, will require stronger centrifugation to remove bubbles. Please use sufficient g-force to ensure that no bubbles are present.

REMOVE ALL BUBBLES

Bubbles can cause optical artefacts as shown in the graph below. Ensure that no bubbles are present in reaction volumes.



KEEP YOUR LAB CLEAN

Please keep your work space clean including all lab equipment like surfaces, pipettes, and tube racks. This will keep the instrument clean and help maintain good results.

KEEP YOUR INSTRUMENT CLEAN

We recommend a routine cleaning of your instrument. To do so follow the cleaning guide in the **Maintenance** section.

PC SETTINGS


Please disable all power saving settings e.g. sleep and hibernate.

EXPERIMENT

This section will teach you everything you need to know to get started with MyGo Mini experiments. You will learn how to create, save, open and close experiments. You will also learn how to set up an experiment including thermal profile, sample information and optical settings. Finally we will show you how to run your new experiment from the software and from the USB drive provided.

CREATE A NEW EXPERIMENT


Below is the MyGo Mini home screen.



Click **New** to create a new experiment.

EXPERIMENT SUMMARY

Once you have created a new experiment the **Experiment Summary** will be displayed as shown below:




By default the experiment name will be “**New Experiment**” with a date and time stamp, which can be edited.

The summary will give you information about the instrument you are running, the settings chosen and any useful notes you choose to add about your experiment.

OPEN AN EXPERIMENT

By selecting **Open** you can open an experiment as shown below:



FILES OF TYPE

The **Files of Type** drop down menu gives you the option to open an experiment normally or as a template that can be used for a new run.



EXPERIMENT FILE

An experiment file will contain all the data and analyses but can not be re-run.

To open as an **Experiment File** select the first option shown on the opposite page.

EXPERIMENT FILE AS TEMPLATE

Opening an experiment file as a template allows you to use the same profile, samples, settings and analyses to re-run a previous experiment, generating new data. When an experiment file is opened as a template, it has the same:

1. **Temperature profile.**
2. **Optical settings.**
3. **Samples and targets.**
4. **Analyses, including settings.**

If you wish to open an **Experiment File as Template** select the second option.

USB RUN

To open a USB run navigate to the USB location using the **Open** dialog window shown to the left. Select **Experiment File** in the **Files of Type** drop down menu and then double click on the USB run.


SAVING AN EXPERIMENT

Once an experiment has completed it can either be automatically saved to a pre-set location on your computer or manually saved by you. The experiment must be saved before the software is closed to ensure that data is not lost.

AUTOMATIC SAVING

When starting a run you will be prompted with the window displayed below.

To automatically save your data when the experiment has finished, firstly click **Browse** and choose the location to save your run to.




Make sure the first checkbox is selected and your experiments will be automatically saved. To prevent the software from showing this dialog again, select the second checkbox. You can change these settings later by clicking **Configuration** and selecting **General Config**.

MANUAL SAVING

Once your experiment has finished you can save your experiment manually by selecting **Save** or **Save As**. If your experiment is unsaved, you will be prompted to choose a location to save your experiment. If your experiment is already saved you can save any changes made to the experiment by selecting **Save**, which will overwrite the previous version.

CLOSING AN EXPERIMENT

To close an experiment select **Close**. If the experiment has not been saved the following dialog box will appear:



SETTING UP A PROFILE

In this section we will cover Hold Times, Programs and setting up a temperature profile.

HOLD TIMES


During thermal cycling protocols hold times should be chosen to allow for the following: thermal equilibration of reaction volumes; completion of biochemical reactions, for example full length strand extension; and optical data acquisition.

Typically, optical data is acquired at the same temperature that strand extension occurs. By default this time is set to 30 seconds. However, this time can be reduced to a minimum of 3 seconds, to save time, or increased to more than 30 seconds in order to increase PCR efficiency and optical sensitivity.

For other hold times during cycling protocols, by default times are set to 10 seconds. However, this time can be reduced to a minimum of 3 seconds, to save time, or increased to more than 10 seconds in order to increase PCR efficiency.

A thermal profile is required to perform an experiment.

Select the profile tab as shown below:



PROGRAMS

To add a program to your experiment click **Add** in the **Programs** pane. Programs can be deleted by selecting them in the list then clicking **Delete**. The program order can also be altered by selecting the program to move then clicking the **Up** and **Down** arrows.

TEMPERATURE PROFILE

Once programs have been added to the experiment they will appear in the **Temperature Profile** pane in the order they are present in the **Programs** window as shown below:



PROGRAM SETTINGS

To alter the settings of a particular program you should first select it, then refer to the right hand pane. In the example above the **Melting** program is selected and can be edited in the right hand pane.

SAMPLES

You will now learn how to set up samples and targets which can be defined as follows:


SAMPLES

A description of the specimen being analysed e.g. "Mouse #5"

TARGETS

A molecular target e.g. "x174", detected with a specified fluorescent reporter e.g. "FAM"

Samples or targets do not need to be defined before starting a run. They can be added to an experiment, whilst it is running, or once it has finished. With a new experiment created, please select the **Samples** tab as shown below:





ADDING SAMPLES AND TARGETS



Samples and targets can be added and removed from the experiment by selecting + and - in the **Samples** and **Targets** pane.




The order of samples and targets can be altered by clicking the **Up** and **Down** arrows in the **Samples** and **Targets** pane.




ASSIGNING SAMPLES TO WELLS

Assign wells by selecting them and clicking **Set**. Samples can also be unassigned by selecting wells you wish to un-assign and clicking **Clear**.



ASSIGNING TARGETS TO WELLS

Targets can be assigned to wells as a Standard, Unknown or Negative by clicking **Std.**, **Unk.**, or **Neg.**, respectively. Assignments can be cleared by clicking **Clear** with the targets and wells you wish to clear selected. For more information about different types of standard controls please contact us on support@mygopcr.com.



DYE COMPATIBILITY


The MyGo Mini comes pre-calibrated for 11 different fluorescent labels, shown in the table below. The **Auto Dye File Generation** software analysis module enables you to calibrate your MyGo Mini for additional fluorophores if required.

Factory Calibrated Dyes	
--------------------------------	--

	ATTO 490LS, CAL 540, CAL 560, FAM, HEX, LC Green, ResoLight, ROX, SYBR Green, SYTO 9, VIC
--	---

SAVING AND OPENING A SAMPLE SETUP

If you wish to save or open sample and target information setup you can select **Save As...** or **Open**, respectively.



Sample and target setups can be saved in the following formats:

CSV Editable .csv files.

PSD A proprietary locked file format.

RDML The Real-time PCR Data Markup Language (RDML) is a structured and universal data standard for exchanging quantitative PCR (qPCR) data.

RUN SETTINGS

The MyGo Mini software provides default optical settings which do not need to be changed for most applications. To view them, select the **Run Settings** tab.


CYCLING ACQUISITIONS

During cycling, by default, optical data will be acquired with settings compatible with a broad range of assays.

MELT ACQUISITIONS

During melting analysis, optical data will be recorded every 500ms. This rate provides relatively compact experimental data file sizes, and supports demanding High Resolution Melting assays.

If you need greater control of how optical data is acquired during your experiment, select “**Use Advanced Settings**”.



ADVANCED SETTINGS

Here, if necessary, you can tailor optical settings to your assay's characteristics.

Integration Time controls the exposure time, in seconds, of the optical sensor inside the MyGo instrument. Brighter reagents require shorter integration times than dimmer reagents.


Longer integration times will provide improved signal to noise with dim fluorescent reporters. Short integration times will reduce the risk of the fluorescence detection system being saturated by very bright fluorescent reporters.

CYCLING ACQUISITIONS - INTEGRATION TIME (S)

Here you can specify the integration time (in seconds) for optical data acquisitions made during thermal cycling.

MELT ACQUISITIONS - INTEGRATION TIME FOR MELT (S)

Here you can specify the integration time (in seconds) for optical data acquisitions made during melting analysis.



DYE CALIBRATIONS

Here is where you can review the current dyes present in your experiment. You have the following options when working with dyes in your MyGo Mini software:

OPEN

Open dye calibrations from a file to be added to this experiment.

SAVE

Save a dye calibration from your experiment to a file on your computer. These can then be used on multiple computers and MyGo Mini instruments.

REMOVE

Remove dye calibrations from this experiment. Factory calibrated dye files can not be removed.

STARTING A RUN FROM THE SOFTWARE

To start a run from the MyGo Mini software select **Start Run**. You will then be presented with the auto save options (unless you have chosen not to be prompted) and then be asked to choose an instrument from the list of **Registered Instruments**. Select an instrument and press **Select** to begin the run.

STARTING A RUN FROM A USB STICK


To start a run using a USB drive select **Start Run**. Then, when you are asked to choose an instrument, select **Start run from USB**. You will then be prompted to find the location of your USB drive. Once selected press **Save** and safely remove your USB drive as normal. The USB drive can now be placed into the instrument and your experiment will start automatically.

LABELLING YOUR TUBES

You can write on the lids of your tubes. Please do not write on the sides of your tubes. Ink can transfer from the sides of tubes to the walls of the wells, and this can affect the performance of your instrument.

LOADING TUBES INTO YOUR MYGO MINI

In order to ensure that the heated lid is balanced, please make sure that the mount contains a tube in each of positions 1, 5, 9, and 13. These positions can be filled with tubes containing reagents, or empty tubes.



ANALYSIS

This section will teach you about different types of analysis available. These can be grouped into the following groups. For information about more analysis options please contact technical support.

QUANTIFICATION

Determine accurate quantities of template or relative expression levels of genes using **Absolute Quantification** or **Relative Quantification**.

GENOTYPING

Use TaqMan probes or High Resolution Melting to perform SNP genotyping using **Automatic Endpoint Genotyping** or **Automatic High Resolution Melt**.

MELTING

Automatically determine the melting temperature of your amplicons using **Automatic Tm Calling**.

RUN FUSION


Combine data from multiple runs on your MyGo Mini for analysis together.



SELECT ANALYSIS TYPE

You can add an analysis type by selecting the “+” button in the bottom left of the software window.

Analysis types are separated into **Automatic** and **Manual** sections. **Automatic** analysis types use advanced automated data processing to provide accurate results with minimal user intervention. **Manual** analysis types enable users to apply traditional analytical methods.



Ticks and crosses indicate which analyses can be applied to the data in your experiment.

AUTOMATIC QUANTIFICATION


POSITIVE / NEGATIVE CALLING

The first step in amplification analysis is to determine which targets have produced a positive amplification curve, and which amplifications have produced a negative amplification curve.

For every amplification curve the software calculates three metrics, these metrics are used to determine which amplifications are positive. These metrics are:


QUALITY

Quality is a measure of how clean an amplification appears. This value will tend to be higher for amplifications with a clear exponential phase and steady baseline, and lower for noisier or flatter curves.




INTENSITY

Intensity is a measure of the size of the exponential phase of the amplification. Note that this may be smaller for amplifications with a negative baseline drift.




PERFORMANCE

Performance is a measure related to the observed efficiency of amplification at later cycles. This value will tend to be lower for inhibited or out-competed reactions, and higher for efficient reactions.



SETTINGS

The software uses a combination of Quality (Q), Performance (P), and Intensity (I) thresholds to determine if an amplification is positive (green zone) or negative (red zone). All values (Q, P and I) must be above the relevant threshold for an amplification to be called positive. The Settings tab enables you to review metrics for amplifications and set appropriate thresholds.



The dots represent individual amplification curves in your experiment. All dots that are within the top right of the graph are called positive and will be subject to normalization determined in **Advanced Settings**.

To switch between “Intensity vs Quality” and “Performance vs Quality” select the Graph Axes drop down menu. All 3 thresholds are used for calling even if they are not currently displayed on the graph.

ARTEFACT FILTERING

Biochemical and physical factors can cause fluorescence levels to change during a run. These are often observed during early cycles. Filtering out these artefacts can improve the accuracy of amplification analysis. Two settings enable you to control the process of filtering out such artefacts.

EXCL. CYC. MIN.

The software will exclude at least this number of cycles from data before performing analysis.

AUTO EXCL. CYC.

By default the software will automatically determine which cycles at the beginning of a run exhibit artefacts, and exclude this number of cycles from data before performing analysis. This feature can be disabled by deselecting this option.

NORMALISATION OPTIONS

Here you can choose the way your data is normalised before display in the amplification graph. By default Full normalisation is applied to enable the discrimination of subtle differences in amplification. Alternative modes are provided to assist data visualisation when troubleshooting challenging assays.

You can choose between the following modes:

	Normalisation Mode		
	None	Background	Full
Background Correction	✗	✓	✓
Normalisation	✗	✗	✓

BACKGROUND CORRECTION

Background fluorescence, including drift, is estimated and removed. There are many potential sources of background fluorescence, e.g. emissions from unquenched probes.

NORMALISATION


Positive amplifications are rescaled to account for intensity variation between wells. Uncorrected intensity can be affected by a number of factors, e.g. the optical properties of disposables. Normalised amplifications are rescaled to a nominal value of 1 relative fluorescent unit, with exact endpoint values a function of curve shape.

RESULTS AS TABLE

The results of amplification analysis are presented in the Results as Table view.

QUANTIFIERS

If you have assigned standards with different known quantities to your experiment, for example a dilution series, you will be able to see your reaction efficiency by selecting **Quantifiers**. Here you will be shown the relationship between Cq and input template quantity, plus a variety of statistics derived from the standards.



RELATIVE QUANTIFICATION

If you are performing relative quantification analysis then select **Relative Quantification** at the top of the **Auto Quant.** pane.

MULTIPLEXING

DYE PAIRS

The MyGo Mini supports multiplexing with the following dye pairs.

Dye Pair	Notes
FAM, HEX	Cost effective probe synthesis
FAM, VIC	e.g. Applied Biosystems TaqMan SNP Genotyping Assays
FAM, CAL560	
FAM, CAL540	
FAM, ATTO490LS	Recommended for challenging applications

FOR BEST RESULTS:

- Ensure that there is no competition between the different PCR in the multiplex reaction.
- Ensure that probes have low background fluorescence, for example by using dual-quenched probes.
- Ensure that probe signal strength is balanced, by using an appropriate concentration of primers and probes.
- Make sure to select the correct dye for each Target.
- Generate assay specific dye files if necessary.

TIPS

Spectral properties of fluorophores can vary as a result of many factors, including stereochemistry, sequence context and buffer chemistry. If you observe spectral crosstalk, generate assay specific dye files to compensate for these effects.

DYE FILE GENERATION

DYE EXPERIMENT


Dye file generation allows you to calibrate your MyGo Mini for new dyes that are not included in the factory-calibrated list. This calibration is based on a PCR amplification using the relevant dye or labelled probe.

DYE CALIBRATION RUN

You will need to perform a real time PCR run in order to generate a dye file. Every well of your MyGo Mini should contain the same reagents. These reagents should produce a single PCR product, and a strong signal for the dye that you are calibrating. Create a sample named after the dye that you are calibrating, and assign that sample to all wells.

AUTO DYE CALIBRATION

Select the **Auto Dye Calibration** option in the Select Analysis Type window. The **Auto Dye Calibration** analysis module will analyse changes in fluorescence during the dye calibration run to create a new dye file.



DYE DATA


Select this tab to view the wells that will be used in the calibration.

AMPLIFICATION

Select this tab to view the cycles that will be used in the calibration.

EXPORT CALIBRATION

Select the **Dye Data** tab and you can choose to export the calibration to a file, or, use it in your current experiment.



Select the Use Wells drop down menu and you will see the list of samples present in the experiment. Select the sample you wish to create a dye from.

USE CALIBRATION IN CURRENT EXPERIMENT

By selecting this option your new dye file will appear in the target list in this experiment. This option will not save the dye file to your computer.


EXPORT CALIBRATION TO FILE

Alternatively you can save this new dye file to your Mac or PC.

GENOTYPING

AUTOMATIC ENDPOINT GENOTYPING

TaqMan genotyping experiments can be analysed automatically using **Automatic Endpoint Genotyping**. Genotypes of samples are determined by the ratio of endpoint fluorescence between two TaqMan probes. Thresholds for the fluorescence ratios are generated automatically as shown below:



AUTOMATIC HIGH RESOLUTION MELTING

Automatic HRM can be performed by using the **Automatic HRM** module as shown on the following page:

DIFFERENCE, NORMALIZED AND MELT


Once opened the analysis module will automatically determine optimised parameters and generate a set of **Difference**, **Normalized** and **Melt** graphs.

CLUSTERING

The difference curves will then be converted into a clustering graph showing clusters of different genotypes.

GENOTYPES


The software will automatically determine **Inferred genotypes** based on the clustering graph. You can define your control Genotypes using the **Genotypes** pane.



MELTING

AUTOMATIC Tm CALLING

Below an **Automatic Tm Calling** module has been added to the experiment and the software has automatically identified and characterised a melting peak.




Tm values will be presented in the table

AUTO SETTINGS

The user has the option to change the target if the experiment contains more than one target. The signal-to-noise ratio setting determines how big a peak should be before it is called. Increasing this value will increase specificity of peak detection. Decreasing this value will increase the sensitivity of peak detection.



PEAKS AND MELTS

The **Peaks** pane will show you the melt peaks with melting point indicators. The **Melts** pane will show you the melting curves of the same reactions.




RUN FUSION

Run Fusion enables you to combine data from multiple runs from your MyGo Mini together for analysis. To fuse multiple experiments together hold down **cmd** on a Mac or **Ctrl** on a PC and select the files you wish to fuse from the open dialog window as shown below.



Select **Open** and then from the next dialog select **Fuse 2 files into one**. You can also open files as separate experiments using **Open 2 Files**.

If you are fusing experiments from multiple file locations you must select the first experiment and then select the “+” from the top left of the open file dialogue window as shown below.




Now hold **cmd/Ctrl** and select the next experiment. Select **Open** to fuse your experiments together.

EXPORT

When you have finished analysing your data in the MyGo software, you can export the results in a variety of ways, from raw data to user defined custom reports. This section will take you through the steps you need in order to do this.

TABLE EXPORT

From panes showing results in tabular format you can export data using the **Export** button.




To export data click the **Export** button, shown on the previous page. Data can be exported in the following formats:

CSV This is an editable data file that can be opened in many spreadsheet applications.

PDF This file format is suitable for archiving, printing and presentations, but cannot be edited.


RDML EXPORT

The software is able to export Cq values and amplification curves from **Auto Quantification** and melt curves from **Automatic Tm Calling** in RDML format. To do this select **Save as**, as if you were saving your experiment. Now from the **Files of Type** drop down menu, select **Real-Time PCR Data Markup Language** as shown below:



GRAPH EXPORT

From panes showing results as graphs, such as amplification and melt curve data can be exported by selecting the button from the below example.



The user can export information as bitmap images (.PNG), Scalable Vector Graphics (.SVG), Character-Separated Variable (.CSV), or Portable Document Format (.PDF).

PNG This is a bitmap image.


SVG A Scalable Vector Graphic is an XML-vector based graphic that can be used where a high resolution scalable graph is required, for example for papers or web design.

CSV This contains information about the graph you are exporting in spreadsheet format.

PDF This contains the graph you are exporting in the Portable Document Format. The graph is in vector format for higher quality.

REPORT GENERATION

Customisable reports can be generated from your experiment by selecting **Experiment Report** under the **Experiment** tab. This provides you with the ability to choose which parts of the data to include in a report. For example, items such as thermal profiles, sample information and various analysis results, e.g. Tm values or quantities can be included, or excluded from reports.




Please select Experiment Report

MAINTENANCE

This section will help you take care of your MyGo Mini. It covers the following areas:

1. CLEANING
2. DISPOSABLES
3. ENVIRONMENTAL CONDITIONS
4. DISASSEMBLY



CLEANING

DAY-TO-DAY

For day-to-day cleaning wipe the external surface of your instrument with a damp, soft, lint-free cloth. Then dry your instrument with another soft, lint-free cloth.

Notes: Avoid abrasive cloths, towels, paper towels, and similar items that might cause damage. Before cleaning your instrument unplug all external power sources, devices, and cables. Don't get moisture into any openings.

DECONTAMINATION

If your instrument needs decontaminating please follow the instructions contained in the MyGo Decontamination Guide.

Please note that for health and safety reasons you must print and complete a physical copy of the decontamination form, and include this with any instrument or lid that is returned. The form must be sent with the instrument, and we can not accept electronic versions of the form. Failure to comply with these guidelines will result in a charge for decontamination.

DEEP CLEANING WELLS

Automated background subtraction processes mean that low levels of fluorescence contamination in the wells of your instrument will not affect system performance. Should one or more wells of your instrument become contaminated with high levels of fluorescent substances, and need cleaning, please contact technical support for guidance.

DISPOSABLES

We recommend the use of MyGo disposables for optimal results. If you wish to use third party disposables please note the following:

- A. Physical dimensions should be the same as the MyGo disposables to ensure that they fit into your instrument, without damaging it.
- B. Caps should seal tubes effectively in order to prevent variability caused by evaporation, and to minimise the risk of contamination with PCR products.
- C. Disposables should have reproducible wall thicknesses, which are thin enough to ensure rapid temperature equilibration, but thick enough to avoid breakages.

In many cases third party disposables will have inferior thermal and optical characteristics, which will reduce the quality of results obtained from the system.

ENVIRONMENTAL CONDITIONS

YOUR WORK SPACE

Your MyGo Mini instrument should be placed on a flat, dry, surface that is not subject to draughts. Do not install your MyGo Mini instrument directly in the flow of air from an air conditioner or fan. Do not install your MyGo Mini instrument in a dusty environment.

PREVENTING CONTAMINATION

Prevent contamination by wearing gloves, using clean tube racks, and filter tips. Make sure tubes are sealed, PCR product is disposed of, and leaks are cleaned immediately.

ENVIRONMENTAL OPERATING CONDITIONS



Humidity

MAX: 80% at +32°C

MIN: 30% at +15 to +32°C



Temperature

+15°C to +32°C



Pressure

0 to 2000 MAMSL

80 to 106Kpa

ENVIRONMENTAL STORING/ TRANSPORTING/PACKING CONDITIONS



Humidity

10% to 95%

No condensation



Temperature

-20°C to +60°C



Pressure

0 to 3000 MAMSL

70 to 106Kpa

DISASSEMBLY

Please note that your MyGo Mini contains no user-serviceable components inside. Any disassembly of your MyGo Mini instrument will void all warranties.

TROUBLESHOOTING

This section will help you troubleshoot your MyGo Mini if you think something is wrong.

FAQ

Here are some frequently asked questions, and the answers.

How do I create a template? All experiments can be used as a template. Click “Open” and then using the “file type” selection in the file dialog to select the “Template” option, then select the file to open as a template.

Where is the threshold for determining Cq values in Auto Quant? Modern methods of determining Cq values are not based on simple thresholds. Modern methods of Cq determination are model based. Auto Quant fits a model of a PCR amplification to the fluorescence data observed. This model fit then enables the estimation of a number of important parameters including Cq values.

Why do some of my amplification curves drop down before rising in Auto Quant?

Correct background removal in Auto Quant relies on correct qualitative calling. Check that appropriate thresholds are set so that positive amplifications are called positive.

Do I need to use ROX as a passive reference in my MyGo instrument?


Some real-time PCR instruments require the use of ROX as a fluorescent passive reference to correct for optical artefacts. Generally, modern instruments do not require the use of passive fluorescent references. The MyGo instruments do not require the use of ROX as a passive reference.

Is it normal for the lid of my instrument to get warm during use?

Yes.

ERROR MESSAGES

Your MyGo Mini will let you know when something is wrong by displaying an error. Most errors are reported as messages in the Status Bar in the MyGo software. Some errors are reported by the instrument display LEDs.



Status Bar.

If you encounter an error, please make sure that the instrument is running in a lab with the required environmental conditions, tubes have been loaded correctly, the lid has been fitted correctly, and all cables are attached correctly and securely. If the error still occurs, you may need to contact technical support with the details described below.

ERROR MESSAGES - DISPLAYED IN STATUS BAR

Message	Action
Instrument will abort because of a major error.	Please use the log button to view the detailed message, and send this, plus a copy of the experiment ppf file, to technical support.
Saturation Warning	Please use the log button to view the detailed message - this will tell you which wells saw a saturated reading. To correct, please reduce the integration time setting, or reduce the quantity/ concentration of dye or probes.
Lid was removed during run - instrument will abort.	The lid must not be removed until the run finishes. See “MyGo Mini Lights and Sounds” for a description of how to confirm the instrument has finished. If running using the network, check for the experiment completed message.

Instrument timed out waiting for heated lid.

Please ensure the lid is fitted correctly, and the ambient temperature is in the required range of +15°C to +32°C.

Could not start run - instrument has encountered a major error.

Please use the log button to view the detailed message, and send this, plus a copy of the experiment ppf file, to technical support.

Could not start run - please remove tubes from previous run.

Remove the tubes, add new tubes, and start the run again. Please note that you must remove the lid and refit it between runs, so that the instrument knows you have removed any tubes from a previous run.

Could not start run - please insert tubes and fit the heated lid.

Remove the lid, check tubes are positioned correctly, and refit lid, making sure it is fully rotated into locked position, and lights and sounds confirm it is fitted.

Could not start run - instrument has an error in heated lid support

Please report error to technical support.

**Could not start run
- instrument has an
error in heated lid.**

Please report error
to technical support.

Received invalid data (X)

Please report error, including
the specific number reported
as X, to technical support.

**Instrument encountered an
error, and had to abort run.**

Please use the log button to
view the detailed message,
and send this, plus a copy
of the experiment ppf file,
to technical support.

ERRORS REPORTED BY THE MYGO MINI DISPLAY LEDs

Some errors are reported by the instrument display LEDs flashing red. If your instrument display LEDs are flashing red, please contact technical support.

LOG FILES

In the unlikely event that your MyGo Mini has a problem, you can produce a log file to send to technical support. For software faults, click the **Configuration** button > **About** tab > click **Save Logs to File**. For hardware faults please provide the .PPF file for the failed run. Sample names etc. can be removed before saving if necessary. If you cannot send the failed run, please send a run performed with empty tubes.

NOTES



www.mygopcr.com

