

## **Specification Sheet**

Specification Number:8015NR-170SRevision:02Date Issued:10-Mar-20

1. Part Number:

8015NR-170S

2. Photograph:



## 3. Product Information:

Description:	SteriWare <sup>®</sup> Narrow Micro Spatula 0.5ml Red - Sterile
Use:	Single use, powder spatula
Material of Construction:	<ul> <li>High Impact Polystyrene (HIPS) – rigid, red</li> <li>Conforms to FDA 21 CFR 177.1640</li> <li>Conforms to EU Regulations 10/2011</li> <li>Conforms to EC Regulations 1935/2004</li> <li>BSE/TSE free</li> </ul>
Method of Construction:	Single piece, injection moulded
Nominal Volume:	0.5ml
Overall Length	170mm
Blade Length:	40mm
Blade Width:	6mm
Nominal Weight:	4g
Moulding & Primary Packing Environment:	Class 100,000 Cleanroom
Individually Bagged?	Yes (heat sealed PE/PET bag)
Method of Sterilisation:	Gamma Irradiation (25 to 45 kGy)
Number of Samplers per Box:	100

Dry and ambient temperature

5 years from the date of manufacture

This device has not been ATEX tested

Recommended Storage Conditions:

Shelf Life:

ATEX:

QMS<sup>o</sup> 9001 : 2015 R E G I S T E R E D Cert No. 14128455

Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet Specification Number: 8015NR-170S Revision: 02 Date Issued: 10-Mar-20

4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to <u>change.control@sampling.com</u>

5. Country of Origin:

Made and packed in the UK.

6. Repeated Use:

7. Approval:

This is a single-use, disposable sampler. Repeated use will lead to failure.

Signed:

Name: Kuldeep Flora Position: Quality Manager Date: 10-Mar-20



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com